



CERTIFICATION by CUSTODIAN of RECORDS  
Tucson Police Department Crime Laboratory

I hereby certify that these documents are a true and correct copy of the  
periodic maintenance\* for the Intoxilyzer 5000 Serial Number  
64-003336 (#15) maintained by the Tucson Police Department  
Crime Laboratory pursuant to the requirements of the Arizona Department  
of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following  
dates. The acceptable accuracy limits, as defined in A.A.C. R9-14-404.A.3,  
for a 0.080AC solution are 0.070AC to 0.090AC.

Date: 2/23/06 Cal 1 0.079 AC Cal 2 0.078 AC

Date: 3/23/06 Cal 1 0.078 AC Cal 2 0.079 AC

Date: N/A Cal 1                      AC Cal 2                      AC

Between the dates of 2/23/06 and 3/23/06  
the instrument was operating properly and accurately.


The instrument was taken out of service for repair on N/A.  
The instrument was returned to service on N/A.

Jennifer L. Coombs  
Jennifer L. Coombs

3/30/06

ADPS QAS Permit # QAS 29886

SUBSCRIBED AND SWORN to before me this 30, day of  
March, 2006.



Kathy A. Kinney  
Notary Public

\*Includes documents specified in A.A.C. R9-14-404.A.6

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the  
information contained in computer storage devices of the Tucson Police Department.

## EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R9-14-404(A)

## ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 5000\*

AGENCY Tucson Police QA SPECIALIST Jennifer Combs 46160  
(Print Name)  
INTOXILYZER SERIAL # 64-CC3336 (#15) LOCATION DUI HQ  
DATE March 23 19 2006 TIME 1615

☒ Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

## DIAGNOSTIC TESTS

☒ DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .229.

☒ Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

☒ Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

☒ Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

☒ Date check. Mode selection switch S11 on or keyboard menu selection "E".

## OPERATIONAL TESTS

☒ Alcohol free subject Test result 0.000 AC

☒ Error recognition logic system functioning  
Invalid test printed

☒ Proper sample recognition system  
Invalid sample printed  
Deficient sample printed

N/A Completeness of sample purge with collection tube\*\*

☒ Calibration standard 0.080 AC

Results: 0.078 AC & 0.079 AC

Instrument operating properly and accurately. YES ☒ NO ☐

COMMENTS:

SIGNATURE

Jennifer L. Combs  
0. N/A alcohol concentration standard collected for subsequent analysis.\*\*

\*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD  
\*\*THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR  
SAMPLE CAPTURE.

DHS/DSLS/Form C117 (Rev. 7-93)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.



3 of 6  
file

# Tucson Police Department Crime Laboratory

By

#15

CMI INC  
INTOXILYZER - ALCOHOL ANALYZER  
AZ MODEL 5000 SN 64-003336  
03/23/06

TEST	BAC VALUE	TIME
AIR BLANK	.000	16:35
CAL. CHECK	.079	16:36
AIR BLANK	.000	16:36

SUBJECT'S NAME

DUI HQ

INSTRUMENT LOCATION

TIME FIRST OBSERVED

46160

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS



INTOXILYZER INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

#15

CMI INC  
INTOXILYZER - ALCOHOL ANALYZER  
AZ MODEL 5000 SN 64-003336  
03/23/06

TEST	BAC VALUE	TIME
AIR BLANK	.000	16:31
CAL. CHECK	.078	16:31
AIR BLANK	.000	16:32

SUBJECT'S NAME

DUI HQ

INSTRUMENT LOCATION

TIME FIRST OBSERVED

46160

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS



INTOXILYZER INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

15

#15

03/23/96  
16:23

SUBJECT'S NAME \_\_\_\_\_

DUI HQ  
INSTRUMENT LOCATION

TIME FIRST OBSERVED

OPERATOR

ADDITIONAL INFORMATION AND OR REMARKS

INTOXILYZER INSTRUMENT PRINTER CARD

© 1986 by CMI INC.



EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R9-14-404(A)

TS4

5 of 6  
file

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 5000\*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 3/23/00 Time 1615  
Intoxilyzer Serial # 64-003336 (#15) Location DUI HQ  
QA Specialist Jennifer Combs 461160  
(Print Name)

- (✓) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made. Standard value: 0.080 AC
- (✓) 2. Turn on the simulator and allow the temperature to reach  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ .
- (✓) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
- (✓) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.
- ✓ (✓) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
- ✓ (✓) 6. Push Start Test button or press enter on keyboard.
- ✓ (✓) 7. Insert card in response to display.
- ✓ (✓) 8. Air blank completed.
- ✓ (✓) 9. Calibration check completed. Test results 0.078 AC & 0.079 AC
- ✓ (✓) 10. Air blank completed.
- ✓ (✓) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.
- (✓) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE

Jennifer L. Combs

\*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev. 7-93)

#15

CMI INC  
INTOXILYZER - ALCOHOL ANALYZER  
AZ MODEL 5000 SN 64-003336  
03/23/06

TEST	BAC VALUE	TIME
AIR BLANK	.000	16:24
*SUBJECT TEST	.174	16:26
AIR BLANK	.000	16:28

\* DEFICIENT SAMPLE - VALUE  
PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

DUI HQ

INSTRUMENT LOCATION

TIME FIRST OBSERVED

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

46160

INTOXILYZER INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

EMI

#15

CMI INC  
INTOXILYZER - ALCOHOL ANALYZER  
AZ MODEL 5000 SN 64-003336  
03/23/06

TEST	BAC VALUE	TIME
AIR BLANK	.000	16:28
INVALID SAMPLE	.XXX	16:28
AIR BLANK	.000	16:28

SUBJECT'S NAME

DUI HQ

INSTRUMENT LOCATION

TIME FIRST OBSERVED

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

Sweet Breath

46160

INTOXILYZER INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

EMI